


MAXIMUM CABLE LENGTH 350 FEET



- 1 CUT OFF AND INSULATE UNUSED CONDUCTORS ON THIS END.
- 2 PART NUMBER INFORMATION TO INCLUDE SPECIFIC DASH NUMBER PER LENGTH AND CURRENT REVISION LEVEL.
- 3 FOLD OUTER BRAID SHIELD OVER OUTER JACKET. WHEN CRIMPING THE MODULAR PLUG, OUTER BRAID MUST MAKE CONTACT WITH THE MODULAR PLUG SHIELD AS SHOWN IN DETAIL A.

REVISIONS									
DESCRIPTION									
DRAWN		ENGR		DATE		LETTER		ECON NO.	
ADDED -29 @ REV. A -XX TO REV. B						B		5000019546	
DCJ	JWR			3/31/14		C		5000022837	
ADDED -26 & -27 @ REV. A -XX TO REV. C						D		5000023615	
DCJ	KMK			7/10/14		E		5000024928	
ADDED -28 @ REV. A -XX TO REV. D									
DCJ	KMK			7/30/14					
ADDED -29 @ REV. A -XX TO REV. E									
DCJ	KMK			9/10/14					
SOURCE/REF DRAWING: 571445-XX									
PROPRIETARY DATA									
The information and designs disclosed herein are the property of MTS Systems Corporation and may not be used, reproduced or disclosed in any form except as granted in writing by MTS Systems Corporation. This restriction excludes information that is in the public domain or was legitimately in the prior possession of the recipient.									
DRAWN DCJ		CHECK JWR		ENGR JWR		MFG			
DATE 10-17-15		DATE 10-20-15		DATE 10-20-15		DATE			
ASSY: PRESSURE XOCR, D PT									
SIZE D		NUMBER 571462-XX		REV E		SHEET 1		OF 1	